



Product Datasheet

Normal Pregnancy Panel

True Human Serum off the Clot

Application:

The in.vent Normal Pregnancy Panels are designed for the use throughout the whole value chain of in-vitro diagnostic test kits, e.g. for R&D and Quality Control Departments. The entire Range of Laboratory Test Results is represented within the Panels. The Samples consist of native Human Bio-Material with no Additives. Titers are based upon Availabilities.

Product No.:	FPPA 4.7.1.1
Number of Samples	30
Volume per Sample	1.0 ml
Source Material	true human serum off the clot
Details given	product number, sample number, matrix, volume, age, gender, measured laboratory result, assay system, minimum shelf life
Diagnoses	<ul style="list-style-type: none">• 10 samples of 1st trimester (1st to 12th week of pregnancy)• 10 samples of 2nd trimester (13th to 24th week of pregnancy)• 10 samples of 3rd trimester (from 25th week of pregnancy)
Materials Safety	All products are for research or manufacturing purposes only and are not for direct therapeutic or diagnostic use in humans or animals.
Storage	≤ -15°C
Minimum Shelf Life	24 months
Manufacturing Quality System	in.vent Diagnostica GmbH operates a quality management system, which complies with the requirements of DIN EN ISO 9001 and DIN EN ISO 13485.
Pricing	Please inquire for your individual quote.

Contact & Ordering Information



Contact for your Order
+49 (0)3302 55 199 250
sales@inventdiagnostica.de

Neuendorfstraße 17 | 16761 Hennigsdorf | Germany

Send us your individual Request:
[inquiry.inventdiagnostica.de](https://www.inventdiagnostica.de/inquiry)



made in germany

www.inventdiagnostica.de