



Product Datasheet

Oestrogen Collective

True Human Serum off the Clot

Application:

in.vent's Collections are designed for the use throughout the whole value chain of in-vitro diagnostic test kits, e.g. for R&D and Quality Control Departements.

Product No.:	PRCO 4.9.1.1.1
Number of Samples	100
Volume	1.0 ml each sample
Matrix	true human serum off the clot
Details given	product number, sample number, volume, matrix, age, gender, measured laboratory result, assay system, minimum shelf life
Parameter	17-beta-Estradiol // range <5pg/ml - >1000pg/ml
Materials Safety	All material of human origin is to be regarded as potentially infectious. Therefore standard precautions are necessary, if handling the material. All products are for research or manufacturing purposes only and are not for direct therapeutic or diagnostic use in humans or animals.
Storage	≤ -15°C
Minimum Shelf Life	24 months
Manufacturing Quality System	in.vent Diagnostica GmbH operates a quality management system, which complies with the requirements of DIN EN ISO 9001 and DIN EN ISO 13485.
Pricing	Please inquire for your individual quote.

Contact & Ordering Information



Contact for your Order
+49 (0)3302 55 199 250
sales@inventdiagnostica.de

Neuendorfstraße 17 | 16761 Hennigsdorf | Germany

Send us your individual Request:
inquiry.inventdiagnostica.de



made in germany

www.inventdiagnostica.de