



Product Datasheet

Pregnancy Panel – Associated Diseases

True Human Serum off the Clot

Application:

The in.vent Panels of Pregnancy – associated diseases are designed for the use throughout the whole value chain of in-vitro diagnostic test kits, e.g. for R&D and Quality Control Departments. The included Diseases are related to Pregnancy (e.g. pregnancy associated diabetes). The entire Range of Laboratory Test Results is represented within the Panels. The Samples consist of native Human Bio-Material with no Additives. Titters are based upon Availabilities.

Product No.:	FPPA 4.7.8.1
Number of Samples	30 (10 samples of each trimester)
Volume per Samples	1.0 ml
Source Material	true human serum off the clot
Content of Panel	for each trimester: <ul style="list-style-type: none"> • 8 x samples abnormal pregnancy (diagnosis and parameters on availability) • 2 samples normal pregnancy
Details given	product number, sample number, matrix, volume, age, gender, week of pregnancy, minimum shelf life, scheme of processing, results for several laboratory parameters on availability
Materials Safety	All material of human origin is to be regarded as potentially infectious. Therefore, standard precautions are necessary, if handling the material. All products are for research or manufacturing purposes only and are not for direct therapeutic or diagnostic use in humans or animals.
Storage	≤ -15°C
Minimum Shelf Life	24 months
Manufacturing Quality System	in.vent Diagnostica GmbH operates a quality management system, which complies with the requirements of DIN EN ISO 9001 and DIN EN ISO 13485.
Pricing	Please inquire for your individual quote.

Contact & Ordering Information



Contact for your Order
+49 (0)3302 55 199 250
sales@inventdiagnostica.de

Neuendorfstraße 17 | 16761 Hennigsdorf | Germany

Send us your individual Request:
inquiry.inventdiagnostica.de



made in germany

www.inventdiagnostica.de