



Product Datasheet

CRP Control

True Human Serum off the Clot

Application:

in.vent Quality Controls are designed to be used in Laboratory Quality Control Settings, in Quality Departments and R&D Units. They are applicable for solving tasks of precision, accuracy or correctness of in-vitro diagnostic assay measurements and analyte components throughout the whole value chain of in-vitro diagnostic test kits or assays.

Product No.:	QC 2.1.14
Source Material	true human serum off the clot
Bottling	0.5 ml each vial
Provided Levels	8 levels in the range of: 0.5 – 100 mg/ml CRP
Stability	12 months at 2-8 °C
Form	lyophilised
Disease State Testing	tested by CE marked test systems and found negative for HIV I/II antibodies, HBs antigen, HCV antibodies
Pre-Samples	available
Delivered Information on Certificate	product number, matrix, processing, homogeneity, target values, reconstitution instruction
Pricing	Please inquire for your individual quote.

Contact & Ordering Information



Contact for your Order
+49 (0)3302 55 199 250
sales@inventdiagnostica.de

Neuendorfstraße 17 | 16761 Hennigsdorf | Germany

Send us your individual Request:
inquiry.inventdiagnostica.de



made in germany

www.inventdiagnostica.de