

Product Datasheet

SARS-CoV-2 Diagnostic sensitivity Panel - positive specimens

Application:

The invent Diagnostic sensitivity Panel - positive specimens is designed for the use throughout the whole value chain of invitro diagnostic test kits, e.g. for R&D and Quality Control Departments. The entire Range of Laboratory Test Results is represented within the Panel. The Samples consist of native Human Biomaterial with no Additives. Titers are based upon Availability.

	DCDA 4.4.0.20.4
Product No.:	DSPA 4.1.9.39.1
Number of samples	300
Volume per sample	0,15 ml
Source Material	true human serum off the clot without additives Human citrate plasma (apheresis)
Details given	product number, volume, matrix, minimum shelf life, result of Laboratory testing (if applicable)
Content	 20x serum samples <21 days after positive PCR 100x plasma samples >21 days after positive PCR without fever during infection 20x plasma samples > 21 days after positive PCR with fever during infection 60x plasma samples SARS-CoV-2 NP IgG low positive 20x plasma samples SARS-CoV-2 NP IgG high positive 40x plasma samples SARS-CoV-2 Spike IgG positive 20x plasma samples SARS-CoV-2 NP IgA positive 20x plasma samples SARS-CoV-2 Spike IgA positive
Materials Safety	All material of human origin is to be regarded as potentially infectious. Therefore, standard precautions are necessary, if handling the material. All products are for research or manufacturing purposes only and are not for direct therapeutic or diagnostic use in humans or animals.
Storage	≤ -18°C
Minimum Shelf Life	24 months
Manufacturing Quality System	invent Diagnostica GmbH operates a quality management system, which complies with the requirements of DIN EN ISO 9001 and DIN EN ISO 13485.
Pricing	Please inquire for your individual quote.

Contact & Ordering Information



Contact for your Order +49 (0)3302 55 199 250 sales@inventdiagnostica.de

Neuendorfstraße 17 | 16761 Hennigsdorf | Germany

Send us your individual Request: inquiry.inventdiagnostica.de

