



Product Datasheet

SARS-CoV-2 Specificity Panel - negative specimens

Application:

The invent SARS-CoV-2 Specificity Panel negative specimens is designed according to FDA requirements for SARS-CoV-2 antibody tests and for the use throughout the whole value chain of in-vitro diagnostic test kits, e.g. for R&D and Quality Control Departments. The entire Range of Laboratory Test Results is represented within the Panel. The Samples consist of native Human Biomaterial with no Additives. Titers are based upon Availability.

Product No.:	DSPA 4.1.9.39.3
Number of samples	400
Source Material	pre-pandemic true normal human serum off the clot
Details given	product number, volume, matrix, minimum shelf life, result of Laboratory testing (if applicable)
Content	400x pre-pandemic samples, collected before 12/2019
Materials Safety	All material of human origin is to be regarded as potentially infectious. Therefore, standard precautions are necessary, if handling the material. All products are for research or manufacturing purposes only and are not for direct therapeutic or diagnostic use in humans or animals.
Storage	≤ -18°C
Minimum Shelf Life	24 months
Manufacturing Quality System	invent Diagnostica GmbH operates a quality management system, which complies with the requirements of DIN EN ISO 9001 and DIN EN ISO 13485.
Pricing	Please inquire for your individual quote.

Contact & Ordering Information



Contact for your Order
+49 (0)3302 55 199 250
sales@inventdiagnostica.de

Neuendorfstraße 17 | 16761 Hennigsdorf | Germany

Send us your individual Request:
inquiry.inventdiagnostica.de



made in germany

www.inventdiagnostica.de